

# ATTTESTATION



ÉMI-TÜV

Attestation No. CA-2029609 on the "testing of the universal electrophysiological biofeedback system" based on the manufacturer's order.

**Applicant:**

QX World Kft.  
1095 Budapest, Tinódi u.1-3" A" épület 4.em. 93.

According to test report No. R-2029609, we attest that the samples below

**Type of QUEX ED and QUEX S are:**

based on our investigations and the submitted documentation

**PASSED**

the following requirements:

2014/35/EU LVD Directive

MSZ EN 60601-1:2017

(EN 60601-1:2006 + A12:2014 + AC:2010 + A1:2013 + AC:2014)

MSZ EN 60601-1:2006/A2:2022

(EN 60601-1:2016/A2:2022)

MSZ EN 60601-2-26:2015

(EN 60601-2-26: 2015

2014/30/EU EMC Direktíve

MSZ EN 60601-1-2:2016

(EN 60601-1-2:2015)

**Validity:** 05th December 2025,

or until changes in laws, regulations or the product.

Budapest, 05th, December 2022.



András Varjú  
Head of Department  
ÉMI-TÜV SÜD Ltd.  
KERMI Department

**Remark:** The test results are related to the tested samples.

You may not publish an extract, abbreviate or abstract of the test report/certificate, and may not be used to advertise the product. The results in the test report only apply to the individual tested samples and the individual test performed.

ÉMI-TÜV SÜD Kft., KERMI Department,  
H-1043 Budapest, Dugonics u. 11., Telefon: +36 1 210-9570